

§ 660.3

21 CFR Ch. I (4–1–16 Edition)

proper name on both the final container and package label, e.g., human, guinea pig.

(2) Name of the test method(s) recommended for the product on the package label and on the final container label when capable of bearing a full label (see § 610.60(a) of this chapter).

(3) A warning on the package label and on the final container label if capable of bearing a full label (see § 610.60(a) of this chapter) indicating that the product and antigen if supplied, shall be handled as if capable of transmitting hepatitis.

(4) If the product is dried, the final container label shall indicate “Reconstitution date: _____” and a statement indicating the period within which the product may be used after reconstitution.

(5) The package shall include a package enclosure providing (i) adequate instructions for use, (ii) a description of all recommended test methods, and (iii) warnings as to possible hazards, including hepatitis, in handling the product and any ancillary reagents and materials accompanying the product.

(d) *Final container.* A final container shall be sufficiently transparent to permit visual inspection of the contents for presence of particulate matter and increased turbidity. The effectiveness of the contents of a final container shall be maintained throughout its dating period.

(e) *Date of manufacture.* The date of manufacture of Antibody to Hepatitis B surface Antigen that has been iodinated with radioactive iodine (¹²⁵I) shall be the day of labeling the antibody with the radionuclide.

(f) *Retention samples.* Each manufacturer shall retain representative samples of the product in accordance with § 600.13 of this chapter except for that which has been iodinated with radioactive iodine. Retention samples of Antibody to Hepatitis B Surface Antigen iodinated with ¹²⁵I shall consist of a minimum of two complete finished packages of each lot of the diagnostic test kit and shall be retained for a period of at least 90 days from the date of manufacture.

[38 FR 32098, Nov. 20, 1973, as amended at 40 FR 29711, July 15, 1975; 46 FR 36134, July 14, 1981; 49 FR 1684, Jan. 13, 1984]

§ 660.3 Reference panel.

A Reference Hepatitis B Surface Antigen Panel shall be obtained from the Food and Drug Administration, Center for Biologics Evaluation and Research, Reagents and Standards Shipping, 10903 New Hampshire Ave., Bldg. 75, Rm. G704, Silver Spring, MD 20993-0002 and shall be used for determining the potency and specificity of Antibody to Hepatitis B Surface Antigen.

[40 FR 29711, July 15, 1975, as amended at 49 FR 23834, June 8, 1984; 55 FR 11013, Mar. 26, 1990; 70 FR 14985, Mar. 24, 2005; 80 FR 18093, Apr. 3, 2015]

§ 660.4 Potency test.

To be satisfactory for release, each filling of Antibody to Hepatitis B Surface Antigen shall be tested against the Reference Hepatitis B Surface Antigen Panel and shall be sufficiently potent to detect the antigen in the appropriate sera of the reference panel by all test methods recommended by the manufacturer in the package insert.

[40 FR 29711, July 15, 1975]

§ 660.5 Specificity.

Each filling of the product shall be specific for antibody to hepatitis B surface antigen, as determined by specificity tests found acceptable by the Director, Center for Biologics Evaluation and Research.

[40 FR 29712, July 15, 1975, as amended at 49 FR 23834, June 8, 1984; 55 FR 11013, Mar. 26, 1990]

§ 660.6 Samples; protocols; official release.

(a) *Samples.* (1) For the purposes of this section, a sample of product not iodinated with ¹²⁵I means a sample from each filling of each lot packaged as for distribution, including all ancillary reagents and materials; and a sample of product iodinated with ¹²⁵I means a sample from each lot of diagnostic test kits in a finished package, including all ancillary reagents and materials.

(2) Unless the Director, Center for Biologics Evaluation and Research, determines that the reliability and consistency of the finished product can be assured with a smaller quantity of